

---

**PATH MEDICAL GmbH**

Landsberger Str. 65  
D-82110 Germering  
Germany  
Tel. +49 / 89 / 80076502  
Email: [info@pathme.de](mailto:info@pathme.de)  
Internet: [www.pathme.de](http://www.pathme.de)



---

## ***Manufacturers Declaration of Conformity***

### ***Australian Therapeutic Goods (Medical Devices) Regulations 2002***

#### ***DECLARATION OF CONFORMITY PROCEDURES***

This is a declaration of conformity made under clause 1.8 of Schedule 3 Part 1 to the Therapeutic Goods (Medical Device) Regulations 2002.

Manufacturers Name: **PATH MEDICAL GmbH**

Business Address: **Landsberger Strasse 65, 82110 Germering, Germany**

Medical Device: ALGO 7i and accessories: ATA Cable, Electrode Cable.  
Classification: Class IIa

GMDN Codes and Term: 35747 Audiometer, auditory evoked response

Scope of application: All products to which the full quality assurance procedures have been applied.

Each kind of medical device to which the Declaration of Conformity Procedure applies, the production quality assurance procedures have also been applied to the device. Each kind of medical device to which the Technical Documentation applies complies with the applicable provision of the essential principles, the classification rules, and these procedures.

#### **Full Quality Management System Certificate:**

EN ISO 13485:2016 Quality System Certificate No 51260- 14-01  
Issued by DEKRA Certification GmbH  
Notified Body Number 0124

ISO 13485:2016 – MDSAP Certificate No 2247210  
Issued by DEKRA Certification B.V.

#### **EC Certificate of Quality Assurance:**

EC Certificate of Quality Assurance; Certificate No: 51260-16-03  
Issued by DEKRA Certification GmbH  
Notified Body Number 0124  
European Medical Device Directives Annex II Certificate

**Standards Applied:**

ISO 13485:2016 – Medical Devices- Quality Management Systems

EN ISO 14971:2019 – Medical Products – application of risk management to medical products

EN 60601-1 Medical Electrical Equipment Part 1: General Requirements for Safety (Ed.: 3.0)

EN 60601-1-2 Medical Electrical Equipment Part 1.2 – General Requirements for Safety – Collateral Standard: Electromagnetic Compatibility – Requirements and tests.

Authorized Signatory:

A handwritten signature in blue ink that reads "Hans Oswald". The signature is written in a cursive style with a large initial 'H'.

Dr. Hans Oswald, General Management

Date: 2022-07-14